

# BackStop+ MiniStop+

## INSTRUCTIONS FOR USE

### Product Description:

MiniStop Plus™ and BackStop Plus™ disposal system with temporary sharps holder.

### Intended Use:

The MiniStop Plus and BackStop Plus are intended for the collection and disposal of blood and other fluids collected during invasive procedures. The devices also include a removable temporary sharps holder for the short-term storage of needles and scalpels.

### Precautions:

- These devices are intended to be used in conjunction with, and not as a substitute for, normal universal blood and sharp injury precautions.
- All needles should be flushed prior to use.

### REUSE PRECAUTION STATEMENT

For single patient use only. Do not reuse, reprocess or resterilize. Reuse, reprocessing or resterilization may compromise the structural integrity of the device and/or lead to device failure which, in turn, may result in patient injury, illness or death. Reuse, reprocessing or resterilization may also create a risk of contamination of the device and/or cause patient infection or cross-infection, including, but not limited to, the transmission of infectious disease(s) from one patient to another. Contamination of the device may lead to injury, illness or death of the patient.

### Instructions for Use:

- 1.Remove the MiniStop Plus or BackStop Plus from package.
- 2.Apply adhesive strip to the bottom of the basin or top of the basin lid (based on customer preference).
- 3.With the adhesive strip attached, place device securely on sterile working area.
- 4.Use device as indicated.
- 5.If necessary, stabilize sharps holder while carefully placing or removing sharps.
- 6.After use, carefully remove temporary sharps holder from basin. **Warning:** Basin may need to be stabilized with free hand.



- 7.Dispose of sharps unit according to designated guidelines for disposing of contaminated medical sharps.
- 8.Place lid on basin.
- 9.Dispose of closed waste basin according to designated guidelines for the disposal of contaminated human waste.

<b>Rx ONLY</b>	Caution: Federal (USA) law restricts this device to sale by or on the order of a physician.
	Single use
	Caution: Consult accompanying documents.
	Do not use if package is damaged

CE 0086



www.merit.com

Merit Medical Systems, Inc.  
1600 West Merit Parkway,  
South Jordan, Utah 84095 U.S.A.  
1-801-253-1600  
U.S.A. Customer Serv. 1-800-356-3748

Merit Medical Ireland Ltd,  
Parkmore Business Park West,  
Galway, Ireland